PRECISION INSTRUMENTS FOR TEST AND MEASUREMENT .

1417 Four-Terminal Capacitance Standard Instruction Manual



Copyright © 2002 IET Labs, Inc.



534 Main Street, Westbury, NY 11590 www.ietlabs.com TEL: (516) 334-5959 • (800) 899-8438 • FAX: (516) 334-5988 1417 Four-Terminal Capacitance Standard Instruction Manual Form 1417-0100-00



www.ietlabs.com 516-334-5959 Fax: 516-334-5988

The material in this manual is for informational purposes only and is subject to change, without notice. QuadTech assumes no responsibility for any error or for consequential damages that may result from the misinterpretation of any procedures in this publication.

Contents

1

J

1

1

]

7

]

I

J

I

I

J

7

]

L

]

L

7

Manual Changes Warranty Specifications Introduction - Section 1 Operation - Section 2 Use With QuadTech Bridges - Section 3 Theory - Section 4 Service and Maintenance - Section 5

WARRANTY

We warrant that this product is free from defects in material and workmanship and, when properly used, will perform in accordance with applicable IET specifications. If within one year after original shipment, it is found not to meet this standard, it will be repaired or, at the option of IET, replaced at no charge when returned to IET. Changes in this product not approved by IET or application of voltages or currents greater than those allowed by the specifications shall void this warranty. IET shall not be liable for any indirect, special, or consequential damages, even if notice has been given to the possibility of such damages.

THIS WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUD-ING BUT NOT LIMITED TO, ANY IMPLIED WARRANTY OF MERCHANTIBILITY OR FITNESS FOR ANY PARTICULAR PURPOSE.



OBSERVE ALL SAFETY RULES WHEN WORKING WITH HIGH VOLTAGES OR LINE VOLTAGES.

Dangerous voltages may be present inside this instrument. Do not open the case Refer servicing to qulified personnel

HIGH VOLTAGES MAY BE PRESENT AT THE TERMINALS OF THIS INSTRUMENT

WHENEVER HAZARDOUS VOLTAGES (> 45 V) ARE USED, TAKE ALL MEASURES TO AVOID ACCIDENTAL CONTACT WITH ANY LIVE COMPONENTS.

USE MAXIMUM INSULATION AND MINIMIZE THE USE OF BARE CONDUCTORS WHEN USING THIS INSTRUMENT.

Use extreme caution when working with bare conductors or bus bars.

WHEN WORKING WITH HIGH VOLTAGES, POST WARNING SIGNS AND KEEP UNREQUIRED PERSONNEL SAFELY AWAY.





DO NOT APPLY ANY VOLTAGES OR CURRENTS TO THE TERMINALS OF THIS INSTRUMENT IN EXCESS OF THE MAXIMUM LIMITS INDICATED ON THE FRONT PANEL OR THE OPERATING GUIDE LABEL.

Instruction Manual Changes

These supplementary pages contain information of improvements or modifications not documented in the main text. All references to GenRad in the manual now apply to QuadTech, Inc.

Page 1 1.1 PURPOSE

<u>Add.</u> The direct-reading accuracy of the 1417 is not sufficient for checking the accuracy of many, more accurate instruments such as the QuadTech Digibridge® impedance meters except at very high values of capacitance. To check these instruments, one must take advantage of the precise ratios of the 1417 which accurately scale the 1 uF value to much higher values up to 1 Farad. Thus the 1 uF value should be calibrated, at each frequency, to get accurate capacitances of higher values. Calibrations at higher values are unnecessary (see paragraph 5.3).

Page 1 1.2 DESCRIPTION

 <u>Add after the CAUTION note</u>. If the instrument to be tested is able to apply dc at any level use a series blocking capacitor to prevent dc from flowing in the 1417 (see paragraph 2.5.3, added to this change section).

Page 3 2.2.1 Four Terminal Connections

 <u>Add to first paragraph.</u> Connection for QuadTech Digibridges are given in paragraph 3.1.1 (added to this change section). It should be noted that theoretically the CURRENT and POTENTIAL terminals may be interchanged.

Page 3 2.2.2 Terminal Impedances

 <u>Add.</u> For most instruments, including the QuadTech Digibridges, the potential connections can tolerate the higher impedances and thus should be connected to the 1417 POTENTIAL terminals. These series impedances cause errors when making 2-terminal measurements which may be reduced by paralleling the CURRENT and POTENTIAL connections (see paragraph 2.8).

Page 3 2.2.3 Connecting Leads

• <u>Add.</u> If possible, the connecting leads should be twisted to reduce mutual inductance. Twist the two (H & L) potential leads together to reduce susceptibility to magnetic fields and twist the two current leads together to reduce the radiation of a magnetic field.

Page 3 2.2.4 Mutual Inductance in Leads

 <u>Add.</u> Mutual inductances in the connecting leads may be greatly reduced by twisting them as discussed in paragraph 2.2.3 added above. Many instruments make a short-circuit calibration correction which can also greatly reduce mutual inductance errors. If such a calibration is made, try to keep the lead configuration as unchanged as possible between calibration and measurement.

Page 3 2.3.1 Accuracy of the 1417

- Add to fourth paragraph. Fortunately, the newer, more precise instruments for measuring high capacitance can apply such low signal levels that this non-linear effect causes only small errors.
- <u>Add to fifth paragraph.</u> It should be remembered that the final application of most instruments that measure high capacitance is to test electrolytic capacitors that generally have broad tolerances. Highly precise instruments do not have to be calibrated to their full accuracy if this is their only end use.

Instruction Manual Changes (continued)

Page 4 2.3.2 Required Accuracy

 <u>Add to last paragraph</u>. Note that often full accuracy is not required particularly if the end use is for testing low accuracy electrolytic capacitors.

Page 5 2.4.1 The 1uF Calibration

<u>Replace complete paragraph</u>. If accuracy better than 0.25% is required, a calibration measurement of the 1 uF value at all three frequencies is required. This can give an accuracy equal to the ratio accuracy plus the 1 uF calibration accuracy at low signal levels. At higher levels, the accuracy based on the 1 uF values should be within 0.1% at 100 or 120 Hz if Ev is held under 5V or at 1 kHz if Ev is less than 20V (refer to paragraph 2.3.3).

The 1uF values of the 1417 may be calibrated directly on a precision bridge or by comparison with a calibrated 1 uF standard, such as the QuadTech 1409Y (a stable silvered-mica unit), that has been calibrated on a precision bridge. Recommended precision bridges are the QuadTech 1615 (1620 system) or the QuadTech 1616 (1621 system). Both of these bridges apply very low voltage at this capacitance value and have accuracy of 0.01% or better. If a calibrated 1 uF standard is used, the instrument being calibrated can usually be used to make the comparison because it will usually have resolution substantially better than its accuracy. This is particularly true of those instruments (such as the QuadTech 1689, 1692 and 1693) that can read differences in percent.

It is recommended that the transformer cores be cyclically demagnetized before calibration to ensure that there is no residual magnetization due to the previous application of dc or a large transient current (refer to paragraph 5.5.1). Also refer to paragraph 5.3 for more information on calibration and adjustment.

Page 6 2.5.1 Magnetization of the Divider Core

 <u>Correction to end of first paragraph</u>. It can be removed by cyclic demagnetization, refer to paragraph 5.5.1.

Page 7 2.5.3 Use of Blocking Capacitors

<u>Add paragraph.</u> DC current should not be applied to the 1417 terminals as it will affect the
magnetization of the transformer divider cores changing their input impedance and thus have an affect on
the capacitance values, particularly at the lower frequencies (refer to paragraph 2.3.1). If there is a
possibility that either the CURRENT or the POTENTIAL circuits of the measuring instrument can apply
even a low-valued dc, blocking capacitors should be placed in series with these connections. These may
be low-leakage electrolytic capacitors and should have capacitances of 200 uF or higher. The
CURRENT circuit of the QuadTech Digibridge impedance meter may apply a few millivolts of dc and
therefore it is suggested that a capacitor be placed in series with the H CURRENT connection to the 1417
in order to achieve the best measurement accuracy.

Page 7 2.6 USE OF AN EXTERNAL STANDARD

- <u>Change first sentence in paragraph one to read.</u> An external standard may be used to get referenced values between those decade values provided, to extend the range or to reduce the effect of inductance error.
- <u>Change paragraph after the formula to read.</u> The larger the external capacitor, the less error (percent) will occur. Using a 10 uF standard will give the same decade values of capacitance as the internal standard (except 1 uF) with less error due to signal level, but calibrating the 10 uF value is quite difficult. If the external capacitor is much smaller than 1 uF, this error becomes very critical.

Page 17 5.3 CALIBRATION PROCEDURE

- Addition to beginning of the first paragraph. The only calibration required is the calibration of the 1 uF setting at each of the test frequencies. These should be well inside the 0.25% tolerance.
- <u>Add.</u> Calibration Requirements. The 1417 direct-reading accuracy can be checked by measuring the 1 uF values at each test frequency. The direct reading values of the other values should be within the specification (0.25% plus the ratio accuracy) if the 1 uF values are well within specifications and thus no calibration of them is required.

For higher accuracy, the calibrated values of the 1 uF setting should be scaled by the appropriate power of ten. This mode uses the very precise and stable accuracy of the internal transformer dividers. These are checked in the manufacturer (and may be tested later, refer to paragraph 5.4.1), but should not change except in the case of gross damage. It makes little sense to measure and record the higher capacitance values because they also depend on the internal capacitor which is less stable than the ratios and has a rather high temperature coefficient. Moreover, there is no known instrument that is adequate for such a calibration! The 1417 should be considered a precise ratio device requiring only the precise calibration of its internal 1 uF standard (and careful use) to get high accuracy. It is preferable to make these precise 1 uF calibrations at the time of use because of possible variations in the internal standard (refer to paragraph 2.4.1)

However, because the effective capacitance of all settings depends somewhat on signal level, particularly at the lower frequencies (refer to paragraphs 2.3.1 and 4.3), it may make sense to make capacitance calibrations for a specific application that uses relatively high signal levels at 100 or 120 Hz.

Page 18 5.4.1 Divider Ratios.

<u>Change second paragraph to read.</u> The ratios should be those for a and b in Table 4-1 to the following limits in ppm of input or in percent of reading.

<u>1/ab</u>	Limit	<u>% Limit</u>
10	16 ppm	.005%
100	5 ppm	.005%
103	1.6 ppm	.005%
104	.75 ppm	.0075%
105	.8 ppm	.024%
106	.6 ppm	.06%

A high precision divider is required. Note that the 1417 could have application as a precision divider.

Page 18 5.5.1 Demagnetization of Divider Cores

- Change last sentence of paragraph one to read. A QuadTech 1311 set for full output on its 30 V range at 50 Hz can give a suitable signal.
- Change step c to read. Set the 1311 to 50 Hz and to its 30 V range and increase the output for maximum.

Vi Statistic



the second of the for the index of the second of the secon

the set of a lot fact that the white the set of the set

Specifications

Capacitance	Ratio	Ratio Accuracy		D Accuracy		Approximate Terminal Impedance	
Value (Internal Standard)	100 & 120 Hz	1 kHz	100 & 120 Hz	1 kHz	ZA (Ω)	ZB (Ω)	(V)
1 μF			±.001	±.001	0.03	0.03	20
10 µF	0.02%	0.04%	±.001	±.001	7.0	15.5	6
100 µF	0.02%	0.04%	±.001	±.001	3.1	6.4	2
1 mF	0.02%	0.06%	±.001	±.002	1.1	2.2	0.8
10 mF	0.03%	0.2%	±.001	±.005	0.37	0.72	0.5
100 mF	0.1%		±.003	*	0.13	0.23	0.25
1 F	0.25%	•	±.01	*	0.04	0.05	0.06

Capacitance: Internal Standard: 1 μ F to 1 F in 7 switch-selected decade values. External Standard: Indicated capacitance, multiplied by C ext/1 μ F.

Capacitance Accuracy, direct-reading: 0.25% plus ratio accuracy (see table) at 100 Hz, 120 Hz, and 1 kHz, 20 to 25° C, with low applied voltage (<1/4 E max) using internal standard and a proper four-terminal measurement. (May also be used as a two-terminal standard, with a D < 1 and a capacitance change from the four-terminal value of < 1/2% up to 1 mF at 120 Hz or less.)

Capacitance Ratio Accuracy: See table.

Dissipation Factor: 0.01 at 100 Hz, 120 Hz and 1 kHz. For D accuracy, see table.

Terminal Impedances: See figure and table (approx values given). Temperature Coefficient: Approximately -140 ppm/°C.

Voltage Characteristic: Approximately +0.3% change from $0_{\rm V}$ to E max (see table) at 100 Hz. Less at higher frequencies.

Mechanical: DIMENSIONS: (wxhxd): 8.5x5.9x5.25 in. (21.5x14.7x 13.2 cm). WEIGHT: 6 lb (2.7 kg) net, 11 lb (5 kg) shipping.



Description 1417 Four-Terminal Capacitance Standard

-EL. ALL L I State I

Catalog Number

1417-9700

(I WINK E)

Introduction-Section 1

1.1 PURPOSE.

The 1417 Four-Terminal Capacitance Standard is a standard used to calibrate bridges and meters that measure high values of capacitance with a 4-terminal connection. It also has some use in the calibration of instruments that make a 2-terminal connection. Most instruments that measure high capacitance do so at 100 Hz, 120 Hz or 1 kHz; the 1417 is compensated to have good direct-reading accuracy at these frequencies. It can be calibrated to be highly precise, making use of its precise capacitance-scaling ability.

1.2 DESCRIPTION.

7

7

T

7

The 1417 uses 2 precision transformers to multiply the value of a capacitance standard by factors of 10 up to 10^6 . It has an internal standard of 1 μ F so that 7 decade values of 1 μ F to 1 F are switch selected. Other values may be obtained by use of an external standard.

This standard does not act as a capacitor at dc and the application of dc can alter the capacitance values.

CAUTION Do not apply dc.

1.3 CONTROLS AND CONNECTIONS.

There are 2 panel knobs driving rotary switches. The CAPACITANCE switch selects the capacitance value as indicated when the internal standard is used. When an external standard is used, the value is that indicated by the C dial, multiplied by $C_{ext}/1 \, \mu F$ (see para 2.6).

The TEST FREQUENCY switch selects proper padding capacitors and resistors to give good direct-reading capacitance accuracy and a D of .01 at 100 Hz, 120 Hz and 1 kHz (see para 2.7 for use at other frequencies). It also has a fourth position that removes the internal $1-\mu$ F standard and connects instead the EXTERNAL STANDARD terminals (see para 2.6).

There are 4 pairs of binding posts on standard 19-mm (3/4-in.) spacing plus 2 ground (case) binding posts. All pairs have terminals marked H and L, and L terminals are all common.

The pairs of terminals on the left are those that should be connected to the measuring instrument. They are labeled POTENTIAL AND CURRENT, common nomenclature for a 4-terminal standard. Generally, these are connected to bridge terminals with similar markings.

The VOLTMETER terminals on the right are directly across the standard capacitor used, whether it is the internal standard or an external standard. Their primary use is to monitor the voltage across the standard so that corrections can be made when precise measurements are required (see para 2.4). These terminals may also be used to add capacitance to the internal standard (see para 2.6).

The EXTERNAL STANDARD terminals are used to connect an external capacitor to obtain other than decade values (see para 2.6). These terminals are switched in the circuit by the TEST FREQUENCY switch.



Figure 1-1. GR 1417 Standard.

Operation-Section 2

The manner in which the 1417 should be used depends on many factors: the accuracy required, the test frequency, the applied voltage, the capacitance values required, and the instrument with which it is used. The operation is very simple in many applications, but can become more complex, particularly when the highest accuracy is required. It is strongly suggested that the theory and calibration techniques used be thoroughly understood to avoid possible errors.

Because different instruments measure under different conditions and have different accuracy specifications, specific instructions are given in Section 3 for each GR instrument that measures high-valued capacitance. These refer to the following paragraphs which describe procedures and precautions that should be used to obtain accurate calibrations for the 1417, so that it may be used to make accurate calibrations of measuring instruments.

2.2 CONNECTIONS TO INSTRUMENTS.

2.2.1 Four-Terminal Connections.

The connections to 4-terminal measuring instruments are made to the H and L CURRENT terminals and H and L POTENTIAL terminals. Usually the terminals of the measuring instrument use a similar notation (although + may be used instead of H and — instead of L) and, in general, these terminals are connected to the corresponding terminals of the 1417, but not always. Section 3, which describes the calibration of specific GR instruments, gives the proper connections for each instrument.

The 1417 also has terminals that tie directly to its case. It is not usually necessary to make any connection to this terminal. However, if strong fields are present, grounding the case may give better results. There is no direct connection between the case and the internal circuitry.

2.2.2 Terminal Impedances.

The 1417 has impedance in series with each of its terminals as shown in Figure 4-4 and Table 4-2. These have no effect on a perfect 4-terminal measurement. However, no instrument makes such a perfect measurement, so that connections should be made in such a way that the larger terminal impedances are connected to the bridge or meter terminals that are most immune to such errors. This is why the various instruments described in Section 3 have different connections. There is a best way to connect to each unit.

2.2.3 Connecting Leads.

The connecting leads used should be short and use reasonably heavy wire to avoid increasing these terminal impedances. However, in many cases test lead sets are supplied for particular instruments. If these lead sets are used also in the application of the instrument, it is desirable to use them for the calibration as well, so that the calibration is made on the instrument as it will be used. If special lead sets are used it must be remembered that they can affect the calibration.

2.2.4 Mutual Inductance In Leads

While a good four-terminal measurement can be immune to self-inductance in the leads, mutual inductance between the current leads and the potential leads can still cause an error, an error which is very important at very high capacitance. The equivalent circuit of figure 4-3 gives the formula

$$Cmeas = \frac{C}{1 \pm \omega^2 CM}$$

The mutual inductance can be positive or negative. This error is critical at very high capacitance, even at 100 Hz, and much more critical at 1 kHz. For example 2.5 nH gives 0.1% error when measuring 1F at 100 Hz or when measuring 10 mF at 1 kHz. The inductance of 2 coaxial, circular, single-turn coils both of 1-cm radius, 1 cm apart, is about 5 nH; it is obvious that great care must be taken when $\omega^2 C$ is large.

Mutual inductance between the current circuit and the potential circuit occurs in the 1417, the connecting leads, and inside the measuring instrument. The mutual inductance inside the 1417 is usually negligible compared to that of the leads. At the highest capacitance values (100 mF and 1 F) it is approx 1 nH. At lower values it causes negligible error at 120 Hz and less than .02% at 1 kHz.

The sections following, on the calibration of specific instruments, note the effect, if any, of the mutual inductances of these instruments and their lead sets.

2.3 ACCURACY.

2.3.1 Accuracy of the 1417.

With no calibration, the accuracy of the 1417 depends on many factors: the accuracy of the initial adjustment, drift in the standard capacitor, its temperature coefficient, the signal level applied, and mutual inductance in its connections. Under reasonable conditions (see para 2.3.3), its uncalibrated or direct-reading accuracy is 0.25%.

The ultimate corrected accuracy is much better. Measurements have been made to \pm .01%. This corrected accuracy is possible because the inductive voltage dividers used to scale the capacitance values (see Section 4) are extremely precise.

This scaling accuracy is the RATIO ACCURACY given in the specifications. This is the accuracy of the ratio between the capacitance value at any setting to the value at 1 μ F, assuming that the signal level on the internal standard remains constant, or that the change caused by a changing

signal level is negligible, or can be precisely determined. To utilize the good ratio accuracy, it is necessary to know the 1- μ F value precisely.

A changing signal level causes a change in the effective value of the standard because of the non-linearity of the input impedance of the voltage dividers (not a non-linearity in their ratios). The effect is largest at low frequencies and negligible at 1 kHz. Typical changes with level are shown in figure 2-2. Methods of determining this level change and making a suitable correction are discussed in para 2.4.2 and 2.4.3.

At 1 kHz, the level effect is negligible (less than .004%), but (see para 2.2.4) mutual inductance effects become noticeable. The RATIO ACCURACY given in the specifications for 1 kHz include the errors caused by mutual inductance inside the 1417, which are generally negligible compared to that of the leads or the measuring device.

2.3.2 Required Accuracy.

In calibration, it is always desirable to have the calibration accuracy of the standard far better than that of the instrument being calibrated. This is not always possible and is never possible at the highest accuracy level for any type of measurement. As a result, the accuracy of the measuring instrument may depend on the accuracy of the standard used to check it.

Suppose, for example, one has a standard whose calibrated value is known (or can be assumed to be) within $\pm A\%$ and a measuring instrument whose accuracy specification is $\pm B\%$. Now we can say:

1. The instrument is within calibration if it reads the calibrated value of the standard to \pm (B-A)%, if B > A, and

2. The instrument may be within calibration if it reads the calibrated value of the standard to $\pm(A+B)\%$.

There is an area of uncertainty of $\pm A\%$ at both + and - tolerance limits of the instrument. If A>B, it is impossible to say whether the instrument is within tolerance or not.

Fortunately, most manufacturers give conservative tolerances to allow for their calibration uncertainties and for aging and environmental effects, so that measurements made in a good environment generally will be within the B-A tolerance, if the instrument is working properly, but only so long as A is not too large. If the instrument is not within the tolerance, but is within the A+B tolerance, it can be readjusted, or if this is not practical, it should be assigned a new specification of $\pm(A+B)$ %.



Figure 2-1. Methods of connection to binding-post terminals.





AVAILABLE INTERCONNECTION ACCESSORIES



DESCRIPTION	CATALOG NO.
Double-plug patch cord, in-line 920-mm (36-in.) long	0274-9860
Double-plug patch cord, right-angle 920-mm (36-in.) long	0274-9880
Shielded double-plug patch cord, 920-mm (36-in.) long	0274-9883
Single-plug patch cord, black, 920-mm (36-in.) long Single-plug patch cord, red, 920-mm (36-in.) long	0274-9468 0274-9492

In many cases, the accuracy of the instrument is far better than required by a specific application, in which case this broader A+B tolerance is of no consequence. For example, if a 0.1% instrument is used to measure ±10% capacitors, errors of a few hundreths percent in the calibration of the 1417 standard are of no importance. It may even make sense to use the simpler calibration procedures that give reduced accuracy.

2.3.3 Direct-Reading Measurements.

The 1417 should be accurate to ±0.25% without corrections if:

1. Proper 4-terminal connections are made (see 2.2.1 and 2.2.2).

2. The TEST FREQUENCY switch is set to a setting (100 Hz, 120 Hz or 1 kHz) corresponding to the frequency of the test signal,

3. The ambient temperature is within the range 23°C ± 5°C and

4. At 100 Hz or 120 Hz, the level at the VOLTMETER terminals is less than 5 V (or the voltage at the CURRENT terminals is less than 0.25% of the value of Emax given in the specifications) and the dividers used in the 1417 have not been left magnitized (see para 5.3) or

5. A 1-kHz test frequency is used and the voltage on the VOLTMETER terminal is 20 V or less (or the voltage at the CURRENT input terminals is less than Emax).

If these conditions are not all met, the appropriate calibration procedures given below should be used to ensure better accuracy. It is good practice to at least make the 1-µF calibration checks to avoid errors due to long-term drift.

2.4 CALIBRATION CORRECTIONS FOR BETTER ACCURACY.

2.4.1 The 1-µF Calibration.

If accuracy better than 0.25% is required, or if the conditions of para 2.3.3 are not met, a calibration measurement at 1 μ F is necessary. Such a calibration will give accuracy within 0.1% for all capacitance values, if the voltage level on the 1417 is kept within the range of items 4 or 5 of para 2.3.3. If at 100 Hz or 120 Hz, the level on the voltmeter terminals, EV, is held constant for all measurements, the percent correction obtained from the 1-µF calibrating measurement can be used for all values and should give a measurement accuracy equal to the specified RATIO ACCURACY. The level of Ev will not affect the 1-kHz values of the 1417 so long as it is below 20 V.

If at 100 or 120 Hz this level (Ey) is not constant, and high accuracy is required, see para 2.4.2 and 2.4.3.

The 1-µF value of the 1417 may be calibrated with either a precision 2-terminal (or 3-terminal) bridge, or a less precise bridge with good resolution and a precision $1-\mu F$ standard.

A recommended precision bridge is the GR 1615 (or 1620 system). A 2-terminal measurement should be made across the VOLTMETER terminals with the TEST FRE-QUENCY switch set to the frequency used. This bridge will apply a very low voltage (EV) to the 1417. Less precise bridges may also be used, such as the GR 1656 or GR 1608, if they are calibrated by measuring a precise 1-µF standard, such as the GR 1409-Y. Both these bridges have .01% (or better) resolution at even decade values. The standard and the 1417 should be measured and the difference between the measurements is the difference between the calibrated value of the standard and the actual value of the 1417. These bridges normally apply a low voltage to the unknown.

This calibrating measurement may also be made with an instrument that one intends to calibrate with the 1417. As above, both a 1- μ F standard and the 1417 are measured and the difference between them (plus the calibration correction of the standard) is used as a correction for the 1417. However, to be meaningful, the measurement device used

must have a resolution substantially better than the accuracy to which it is to be calibrated.

2.4.2 Approximate Level Corrections.

If at low frequencies (100 Hz or 120 Hz) the voltage level at the VOLTMETER terminal, E_V , is not very low or is not constant at various 1417 capacitance settings, a level correction should be made for best accuracy. The method for making these corrections is given in para 2.4.3. However, for those GR instruments discussed in Section 3, an approximate correction, accurate to 0.1%, can be made without making additional calibrating measurements on the 1417.

The procedure is as follows:

a. Measure the low-level $1-\mu F$ value of the 1417 at the desired test frequency as in para 2.4.1 and note the percent deviation from the nominal value.

b. From the tables of Section 3, note the value of ${\rm E_V}$ for each specific range of the instrument used and 1417 setting. Then, from the typical calibration curve of Figure 2-2, determine the approximate percent capacitance change.

c. Apply the corrections of a and b above to the measurements made on the higher 1417 settings.

Better accuracy can be obtained if a capacitance vs level curve is made for the specific unit used.

2.4.3 The Capacitance vs Level Corrections.

To obtain the best accuracy at low frequencies, it is necessary to make a precise capacitance-vs-level correction chart for the specific 1417 used, to measure the voltage at the VOLTMETER terminals at each setting of the 1417 during use, and to apply the previously determined correction. The curves should be similar to the typical correction curves of Figure 2-2. (This is not necessary at 1 kHz). It is preferable to demagnetize the dividers of the 1417 before these capacitance-vs-level corrections are determined (see para 5.3).

A precision bridge is not necessary to make this correction curve, but the bridge should have good resolution and must be able to apply 20 V to the $1-\mu$ F value or at least as much voltage as will be required in use. The bridge may be 2-terminal because the connection of the 1417 is to the VOLTMETER terminals.

The GR 1615 Precision Capacitance Bridge is not suitable for these calibration measurements at low frequencies because in its normal connection it applies very little voltage to a $1-\mu F$ capacitor and in its reverse connection the detector mismatch results in poor sensitivity.

The GR 1656 Impedance Bridge may be used, and, with a GR 1311 oscillator, allows measurements to 20 V at 100 and 120 Hz. To obtain precision at low-levels, the recommended connection of Figure 2-3 should be used, the operator, bench, and all nearby equipment should be grounded, and the operator should not touch the metallic surfaces of the 1656 during balance. While this connection leaves the case of the 1656 ungrounded, the 1656 is battery powered, so that hum from external sources only can affect the measurement. This connection grounds the 1656 HIGH UNKNOWN terminal and the L VOLTMETER terminal of the 1417, so that a voltmeter with a grounded low terminal can be used. The voltmeter should be disconnected from the circuit for the final balance, but the voltage should be measured when the 1656 is fairly well balanced, because the voltage on the 1417 depends on bridge adjustments. With this method, resolution to .01% down to 0.2 V, is possible if care is used.

The 1656 should be set to the 100 nF MULTIPLIER setting, so that readings will be made at full scale to obtain .01% readout resolution. All readings should be made using a 9 as the first (left most) digit of the readout. This allows balance up to 9XXX = 10110, which is more than adequate and minimizes the effect of errors in the main bridge adjustment.

Since the 1656 is only 0.1% accurate and because the connection may introduce a small, but constant error, the resulting calibration curve is only precise for relative values. The precise value should be obtained at a specific level (preferably a low level) by calibrating the 1656 in this connection with a precision standard or by a measurement on a precision bridge, as described in para 2.4.1.

2.5 PRECAUTIONS.

2.5.1 Magnetization of the Divider Core.

CAUTION Do not apply dc to the 1417.

Dc applied to the 1417 can make a severe permanent change in the inductance of the transformer dividers, which can affect the low-frequency calibrated value. A

CAUTION

A 500- μ F capacitor must always be connected between the H terminal of the 1417 and the current terminal of any bridge connected to it. This protects the 1417 from possible damage.



Figure 2-3. Connection diagram for 1417 calibration checks.

noticeable change can be made by as little as 0.1 ampereturn and, since there are 4000 turns on one divider, it could be affected by as little as $25 \,\mu$ A. The result is a reduction in the inductance of the cores and a reduction in the effective value of the 1417. This effect is a result of residual magnetization of the cores. It can be removed by cyclic demagnetization, see para 5.3.

A residual magnetization can also result from switching large ac signals on and off. If the signal level, E_V , is above 10 V at 100 or 120 Hz, it is preferable to reduce the drive signal slowly before connecting or disconnecting the 1417.

2.5.2 Maximum Input Voltage.

The maximum input voltage is given in the specifications. This limit is based on a maximum power of 2 W, a current of 2 A and a voltage of 20 V, whichever is limiting, at 100 Hz. These limits are adequate for testing all GR instruments.

Aside from possible overheating, the reason for these limits is the two 1.5-W Zener diodes (in series opposing) that shunt the VOLTMETER terminals. These clamp the output to a safe value to protect the operator and the instrument. If the voltage at these terminals exceeds about 27 V rms, these diodes will clip and cause measurement errors. A maximum of 20 V rms is recommended. If the 27-V limit is exceeded and the source can deliver high power, the protecting Zener diodes may be destroyed, either shorting, making the 1417 unusable, or opening, and removing the protection feature. Usually they will short unless extreme power is applied.

2.6 USE OF AN EXTERNAL STANDARD.

An external standard may be used to get referenced values between those decade values provided or to extend the range. This can be accomplished in two ways: by using the external standard by itself or by adding it to the internal $1-\mu$ F standard.

When an external standard is connected across the EXT-ERNAL STANDARD terminals and the TEST FREQUENCY switch is set to that position, the 4-terminal capacitance is ideally that indicated on the CAPACITANCE switch multiplied by $C_{ext}/1\mu$ F. This value is subject to the errors discussed in para 4.3.

The low frequency error of the effective value of the EXT CAP is

$$C'_{ext} = C_{ext} \left(1 - \frac{1}{\omega^2 L C_{ext}} \right)$$

The larger the external capacitance, the less (percent) error will occur. This error can be quite large at low frequencies if C_{ext} is substantially less than 1 μ F. This error can be corrected for by use of the procedures of para 2.4 If C_{ext} is low, these corrections will be critical.

When using an external capacitor connected this way, the D value will be the D of the external capacitor plus a small quantity. The D should be determined by a measurement at the $1-\mu F$ position and it may be set to a desired value by adding resistance in series with the external standard.

When an external standard is connected to the VOLT-METER terminals, it is added (in parallel) with the internal standard and thus increases all values. Ideally, the measured capacitance is the value indicated by the CAPACITANCE switch multiplied by $(1 + C_{ext}/1 \mu F)$. This connection has one advantage in that the internal padding capacitors used to compensate for the low-frequency inductance error (para 4.3) are still effective (if the proper TEST FRE-QUENCY position is used). The effective value is

$$C' = 1 \,\mu F + C_{ext} + Cp - 1/\omega^2 L$$

where Cp is the padding capacitor that cancels the inductance term. Moreover, the larger C_{ext} is, the less the (percent) error caused by this term. However, for precise measurements this error should be determined as before para 2.4.

The D values will be approximately

(.01) $1\mu F/(1\mu F + C_{ext})$.

For precise D calibrations, this value should also be determined by a measurement at the $1-\mu F$ setting.

2.7 APPLICATION OF OTHER FREQUENCIES.

The value of the GR1417 depends somewhat on frequency (see Figure 2-4). It is padded to give good directreading accuracy at 100 Hz, 120 Hz and 1 kHz, but may be used at other frequencies. For frequencies between 100 Hz and 1 kHz, the full accuracy of the 1417 may be realized by measuring the 1- μ F value of the 1417 at the desired frequency in the 1 -kHz position of the TEST FREQUENCY switch and either noting the error for future use or increasing the value to 1 μ F, by adding padding capacitors at the VOLT-METER terminals.

Below 100 Hz, the low-frequency error (see para 4.3) increases and, while corrections or padding should be used, level effects will be increased so that full corrected accuracy becomes more difficult to maintain.



Figure 2-4. Typical capacitance change with frequency. Note corrections are made at 100 Hz and 120 Hz.

Above 1 kHz, the low-frequency error is negligible but the effect of mutual inductance (para 4.3) is increased and eventually the divider ratios deteriorate. The increased error (both ratio and direct reading) should be less than $\pm (.02\%)$ (f_{kHz})², if the mutual inductance external to the 1417 is negligible. The 1-kHz TEST FREQUENCY position should be used.

At frequencies other than 100, 120 and 1000 Hz the D readings will not be .01. They will be approximately (.01) f/f_s , where f is the test frequency and fs the setting of the TEST FREQUENCY switch. For accurate D values, an external series capacitor and resistor should be used which, on the 1 μ F position, give a calibrated value of 1 μ F and a D of .01.

2.8 TWO-TERMINAL MEASUREMENTS.

The lower values of the 1417 may be used as a 2-terminal standard at low frequencies. The D is increased because of the terminal resistances, and the C value is changed slightly because of terminal inductance and the variation in the α and β ratios (effective terminal capacitance see para 4.4). All those effects are reduced if the CURRENT and POTENTIAL terminals are connected in parallel. Table 2-1 shows approximate D values and the change in capacitance from its 4-terminal value for both this parallel connection and for connections to the current terminals only at 120 Hz.

Table 2-1

TWO-TERMINAL MEASUREMENTS*

Current 1417 Terminals Setting △C(%)		D	Current & Potent. Term. in Parallel △C(%)	D	
1 μF	0	.01	0	.01	
10 µF	+0.3	.063	+0.2	.045	
100 µF	+0.8	0.24	+0.4	0.17	
1 mF	-1.4	0.85	-0.15	0.56	
10 mF	+2.5	2.85	+1.5	1.92	

*Typical C and D values.

Use with GR Bridges-Section 3

3.1 USE WITH 1617 BRIDGE.

The 1617 Capacitance Bridge is a 1% manual unit whose accuracy depends on its dial calibration (an internal adjustment), its standard capacitor, and its range resistors. The dial calibration and standard can be checked with a precision capacitance decade box (see 1617 Instruction Manual), but a standard of high capacitance, such as the 1417, is required to check the higher range resistors. The 1617 dial is logarithmic and equally accurate at readings of 1 or 10. Therefore, the ranges can be checked at either point, or both.

Because the 1617 accuracy is only 1%, and the applied power is low (if the recommended test voltages are used), the 0.25% direct-reading accuracy of the 1417 is usually adequate, but a single calibration of the 1417 at 1- μ F (see para 2.4.1) is nevertheless suggested as a precaution. The voltage, E_V, reaches 3 V on two measurements, which would mean a capacitance change of +0.1%, but this should be negligible compared to the ability to read the 1617 dial.

The 1617 makes a 4-terminal measurement, but errors occur if the lead impedances or terminal impedances of the 1417 are sufficiently high. To keep these errors at a minimum, different connections are used on different ranges (see Tables 3-1 and 3-2). The remaining lead errors affect D on some ranges, giving a repeatable error on two lower ranges and a less precise error in three higher ranges. To check the D accuracy on the higher ranges, measurements should be made at 10 on the dial.

Table 3-1 also lists the recommended 1617 test voltages (GEN LEVEL switch). When measuring high capacitance,

the 1617 is subject to error caused by hum pickup. This error can and should be removed by making two measurements with the METER switch set to GEN NORM and GEN REV and taking the average readings.

The 4-terminal lead set (P/N 1617-2210) should be used (with the shorting links on the 1617 unknown terminals disconnected.) At very high values, the mutual inductance between the leads is important. The potential leads (the two inner leads) should be tightly twisted to reduce this error (see para 2.2). If this precaution is not taken, the error can be more than 10% at 1 F.

Table 3-2 TEST CONNECTIONS

1417 Connections				
"A"	"B"			
HPOTENTIAL	L CURRENT			
H CURRENT	L POTENTIAL			
L CURRENT	H POTENTIAL			
L POTENTIAL	H CURRENT			
	1417 Connect "A" H POTENTIAL H CURRENT L CURRENT L POTENTIAL			

3.2 USE WITH 1657 Digibridge.

The accuracy of the 1657 depends mainly on three precision resistors and a crystal oscillator. These can most easily be checked by resistance measurements of 10 Ω , 1 k Ω and 100 k Ω , and one precise capacitance measurement. However, if one wants assurance that it measures high capacitance values accurately also, one may use the 1417.

		an provinsi siya kiya na ma dana pala ma	Tab	le 3-1				
			CALIBRATIC	ON WITH 14	17			
1417 Setting	1617 Multiplier	Connection	Gen Level(V)	Nom C Read.	C Tol.	D Nominal	D Tol.	EV (120 Hz)
1 μF	100 nF	A or B	2.0	10	±1%	.01	±.001	0.24
1 µF	1 µF	A or B	2.0	1	± 1%	.01	±.001	1.45
10 µF	1 µF	А	2.0	10	± 1%	.008	±.001	0.77
10 µF	10 µF	A	0.5	1	± 1%	.008	±.001	1.2
100 µF	10 µF	A	0.5	10	± 1%	.009	±.001	0.64
100 µF	100 µF	А	0.2	1	± 1%	.009	±.001	1.4
1 mF	100 µF	В	0.2	10	± 1%	.01	±.001	0.75
1 mF	.1 mF	В	0.2	1	± 1%			3.0
10 mF	1 mF	В	0.2	10	± 1%	.01	±.0011	1.55
10 mF	10 mF	В	0.2	1	± 1%			3.1
100 mF	10 mF	В	0.2	10	± 1%	.01	±.002	1.50
100 mF	100 mF	В	0.2	1	± 2%	-		1.55
1 F	100 mF	В	0.5	10	± 2%	.01	±.011	1.0

NOTES (1) Use 1417 frequency setting corresponding to test frequency.

(2) Make two measurements with 1617 input direct and reversed and take average.

(3) Twist leads at C values >1 mF.

The 1657 is 0.2% accurate at the lower values, so that the 1417 should be calibrated before it is used. However, because the 1657 applies low power (see Table 3-3) a single, low-level calibration (see para 2.4.1) for each frequency is all that is required.

Connections between the 1657 and the 1417 should be made as given in Table 3-4. The 1657 I+ connection may have a slight dc offset voltage. A large capacitor (500 μ F) should be put in series with the connection.

A test lead set should be used. The mutual inductance of the 1657 gives negligible error compared to its specifications, but care should be used in the lead configuration when measuring 1 mF at 1 kHz.

Note that the maximum reading is 999.99 mF at 1 kHz and 99.999 mF at 120 Hz. It would be preferable to use an external standard with the 1417, of say 0.9 μ F to check these range extremes.

			Table	3-3			
		1	657 ACC	URACY			
1417	120	Hz		1 kH	z		Ev
Setting	∆ C%	D	Range	∆C(%)	D	Range	(120-Hz)
1 µF	0.2	.001	2	0.2	.001	2	0.15
10 µF	0.2	.001	2	0.2	.001	1	.07
100 µF	0.2	.001	1	0.2	.001	1	1.25
1 mF	0.2	.001	1	1*	.01	1	0.64
10 mF	1	.005	1	Margin (Marin	-	1	Low
100 mF	10 **	.05	1	-	-	1	Low
100 1111	10	.00					6000

*Goes to 999.99 µF. **Goes to 99.999 mF

Table 3-4 INTERCONNECTION

1657	1417
P- (rear left)	L POTENTIAL
P+ (rear right)	H POTENTIAL
I- (front left)	LCURRENT
I+ (front right)	H CURRENT

3.3 USE WITH 1683 BRIDGE

The GR 1683 Automatic RLC Bridge measures at both 120 Hz and 1 kHz to a base accuracy of 0.1% (see Table 3-5) so that calibration to full accuracy requires careful calibration and use of the 1417.

CAUTION

GR 1683 applies 2.2 V at 1 A; danger of residual magnetization (see para 2.5.1.)

At 120 Hz, a voltage-vs-level curve should be used to obtain full accuracy (see para 2.4.2). At 1 mF full output of the 1683 can give 25.5 V on the 1417 VOLTMETER terminals, which is close to the Zener-diode clipping point. It is suggested that the GR 1683 OSC LEVEL be set to 1 V for this measurement. Because of the high level supplied by this bridge, it is recommended that the 1417 be connected and disconnected and all range changes made with the 1683 OSC LEVEL set to 0.

At 1 kHz there is negligible level error but the above precaution is recommended to avoid changes in the 1417 calibration due to transients.

The measurement connections between the 1683 and 1417 are given in Table 3-7. The mutual inductance error

caused by the 1683 is negligible compared to its specifications, if the current leads are twisted together and the potential leads are twisted together.

a.	16	Table 3-5 83 ACCUR	ACY*	
1417	120	Hz	1 k	Hz
Setting	△C(%)	D	∆C(%)	D
1 µF	0.11	.0011	0.11	.0011
10 µF	0.11	.0011	0.11	.0011
100 µF	0.11	.0011	0.11	.0011
1 mF	0.11	.0011	1.1	.01
10 mF	1.1	.01	6	0.1
100 mF	6	0.1	()maxim	

*1683 C reading of 10000 and D = .01

Table 3-6

TYPICAL VOLTAGES AT VOLTMETER TERMINALS 1683 OSCILLATOR SETTING (120 Hz)

Setting	Full	1 V
1 μF	2.1	1
10 µF	5.3	2.5
100 µF	14.7	6.9
1 mF	25.5	12
10 mF	17,8	8.4
100 mF	7.6	3.6

Table 3	1-7
BRIDGE-STANDAR	D CONNECTIONS
1683	1417
+CURRENT	CURRENT H
-CURRENT	CURRENT L
+POTENTIAL	POTENTIAL H
POTENTIAL	POTENTIALI

3.4 WITH 1685/1686 INSTRUMENTS.

The 1685 Digital Impedance Meter and the 1686-A and 1686-9000 Digital Capacitance Meters are very similar in operation and specifications. The 1685 measures resistance and inductance as well as capacitance and its ratio and accuracy can be most easily checked by measurements of precision resistors. However, a 1417 is useful for checking its operation for high-capacitance measurements. The 1686 only measures capacitance, so a standard such as the 1417 is necessary to check its high-capacitance ratio errors.

The 1685 and 1686 have the same specifications for capacitance, but the 1686 applies slightly more power on some ranges (see Tables 3-8 and 3-9). The 1686-A has slightly improved specifications compared to the 1686-9000 and has an additional range up to 200 mF. Each has rearpanel level control that should be set full on (cw). All these instruments will make a measurement at 120 Hz up to 5 times the full-scale value on the highest range. The error "E" will appear in the readouts but the measurement is valid, if not too accurate (see Tables 3-8 and 3-9).

The connections are the same for all three instruments and are given in Table 3-10. The test lead sets P/N 1685-9600 for 1685, P/N 1686-9602 for 1686 should be used. These instruments with their lead sets have approx. $0.2 \,\mu$ H mutual inductance between the current and potential pairs which

Į

causes an error on the highest ranges (see Table 3-11 and para 4.3). This error increases the measured value and should be subtracted from all measurements to obtain the best accuracy.

		Table	3-8			
	1685	/1686-900	120 Hz			
1417	120 Hz		lz 1 kHz		1686 EV	1685 EV
Setting	26(70)	U	DC(70)	U	Approx	Approx
1 μF	±0.35	±.0019	±0.12	±.0019	1	0.95
10 µF	±0.35	±.0019	±0.12	±.0019	3.5	3.3
100 µF	±0.35	±.0019	±0.36	±.0057	7.7	7.7
1 mF	±0.7	±.0038	±1.2**	±.019	1.7	1.7
10 mF	±1.4	±.0176			0.9	0.9
100 mE	61				04	04

t Overrange.

*At half full scale and D = .01. **Range not recommended.

		1686-A ACCU	RACYT		
1417 Setting	120 H ∆C(%)	D	1 kł	iz D	E _V (120-Hz)
1 µF	±0.35	±.0019	±0.12	±.0019	1
10 µF	±0.35	±.0019	±0.12	±.0019	3.5
100 µF	±0.35	±.0019	±0.36	±.0057	7.7
1 mF	±0.7	±.0038	±1.2	±.019	1.7
10 mF	±1.05	±.0176			0.9
100 mF	±3.4	±0.11	-	-	0.4
1 F	28*				0.1

Table 3-9

• Overrange

tAt half full scale and D = .01.

Table 3-10

	CONNECTIONS	
168	5-1686 Cable	1417
[+	RED	H CURRENT
P+	RED, WHITE SLEEVE	HPOTENTIAL
P	BLACK, WHITE SLEEVE	L POTENTIAL
1-	BLACK	L CURRENT

Table 3-11 MUTUAL INDUCTANCE ERROR-(APPROX.)

1417 Setting	120 H	1 kHz
1 μF	0	0
10 µF	0	+.01%
100 µF	0	+.08%
1 mF	+.01%	+.8%
10 mF	+0.11%	+ 8%
100 mF	+1.1%	
1 F	+11%	

3.5 WITH 2230 SYSTEMS.

The bridge of the 2230 Component Network System actually measures C parallel and G parallel, but the computer in the system converts the measurement data to give a display of C series and D. It measures capacitors at 120 Hz and 1 kHz to good accuracy (see Table 3-12) so that the correction procedures of para 2.4.2 or 2.4.3 should be used. The mutual inductance of the leads of this bridge does not affect the accuracy because of its limited capacitance range. The connections are given in Table 3-13.

One precaution should be considered. The I+ connection is not decoupled from the signal source. This source should have a very low offset voltage, but if not adjusted properly it can supply direct current to the 1417. If the measurements are in error, a large capacitor should be placed in series with this connection to determine if this is cause of the inaccuracy.

	Table	3-12			
2230 AUTOMATIC	TEST	SYSTEM,	CGRL	BRIDGE	

	, 1	, 1	kHz				
Cap. Value	Range	Accurac C	D	Εv	Range	Accurac C	D
1 µF	5	±.125%	±.001	7	6	±.12%	±.001
10 µF	6	±.12%	±.001	3.1	7	±.12%	±.001
100 µF	7	2.12%	:.001	8.4	8	2.14%	±.002
1 mF	8	:.14%	:.002	1.6			-

TABLE 3-13 INTERCONNECTIONS

2230	1417
1+	H CURRENT
P+	H POTENTIAL
P-	L POTENTIAL
1-	L CURRENT

3.6 USE WITH QuadTech Digibridges 3.6.1 General

The 1417 Four-Terminal Capacitance Standard is useful for checking the accuracy of the QuadTech low frequency Digibridges at high capacitances. These instruments are calibrated using precision resistors, but the 1417 can give assurance that they measure high capacitance accurately. These instruments have different specified accuracy and different capacitance ranges so that detailed information is given for each in the paragraphs that follow. All of them have accuracy better than the 0.25% direct-reading accuracy of the 1417 and therefore 1 uF calibrations of the 1417 are required and some care necessary to realize its high ratio accuracy (refer to paragraphs 2.4.2 and 5.3). In some cases the specified Digibridge accuracy is as good as the 1417 ratio accuracy so that a true accuracy verification is not possible but nevertheless, tests with a 1417 gives added confidence.

The specified accuracy given for all instruments are for their SLOW measurement rates if there is a choice of test rates. Greater repeatability can be achieved by averaging several measurements either manually or automatically if this feature is available.

3.6.2 Connections

Four-terminal connections should be made to the 1417 as shown in Table 3-14 below using the 1657-9600 Extender Cable (1657, 1658, 1659, 1689, 1692) or the 1689-9602 Extender Cable (1689M, 1693). The Digibridge current terminals may have a slight dc voltage so that a large capacitor, > 200 uF, should be connected in series with the H CURRENT terminal to keep dc from flowing in the precision transformers (this may be an electrolytic capacitor).

Table 3-14 Digibridge Test Connections

		Digibrid	lge Connection		
1417	Cable Plug	1657	1659, 1689	Test Fixture	
Terminal	Color Code	1658	1692, 1693	Connection	
H POTENTIAL	Red-white	P+	PH	right rear	
L POTENTIAL	Black-White	P-	PL	left rear	
H CURRENT*	Red-Red	I+	IH	right front	
L CURRENT	Black-Black	I-	IL	left front	
CASE	Green	Ground	Ground	threaded holes	

*Add series capacitor

3.6.3 Zeroing and Minimizing Mutual Inductance

Mutual inductance between the test leads can cause an error, particularly at 1 kHz (refer to paragraph 2.2.4), even though the position of the terminals on the 1417 minimizes this effect. It is also suggested that the free ends of the test cable be twisted to avoid coupling between the current and potential circuits. Twist the two POTENTIAL leads together and twist the two CURRENT leads together. Many instruments have short-circuit calibration capability that should be used to remove this effect. The suggested procedure is as follows:

a. Set the Digibridge to the test frequency to be used.
b. Connect the 1417 as in Table 3-14 except plug the high potential plug (red-white) into the low potential plug (black-white) which is plugged in the L POTENTIAL terminal of the 1417.

c. Set the 1417 CAPACITANCE switch to 1F.

d. Make the SHORT calibration of the test instrument.
e. Reconnect the high potential plug into the H
POTENTIAL terminal being careful to move the leads as little as possible.

f. Reset to desired capacitance value.

3.6.4 Use with the QuadTech 1657 or 1658 Digibridges.

These two instruments both have three ranges which are usually checked by three calibrated resistors (10 ohms, 1 kohms, and 100 kohms). Their capacitance ranges extend to 99999 uF at 100 or 120 Hz and 999.99 uF at 1 kHz. This means that 100 mF at 120 Hz and 1000 uF at 1 kHz may only be measured if they are lower than the nominal value. A chart of the C and D accuracy of the 1657 vs capacitance is given in Table 3-15. Divide the accuracy values by two for the 1658 using its SLOW measurement rate.

Both of these instruments can apply enough current to give a value for Ev of over 1 V at 100 uF (refer to Figure 2-2) but both instruments should still measure well within their accuracy (relative to the 1 uF calibrated value at each frequency).

Neither of these instruments makes a SHORT calibration so that care should be used to keep the mutual inductance at a minimum (refer to paragraph 3.1.2).

	Table 3	-15
1657	Specified	Accuracy

1417	100 or 120 Hz			1 k	Hz	
Setting	С	D	Range	C	D	Range
1 uF	0.2%	.001	2	0.2%	.001	2
10 uF	0.2%	.001	2	0.2%	.001	1
100 uF	0.2%	.001	1	0.2%	.001	1
1 mF	0.2%	.001	1	1%*	.01	1
10 mF	1%	.005	1	_	-	
100 mF	10%*	.05	1 .		-	÷ *
			ž.			

* At range limit: 999.99 uF at 1 kHz, 99999 uF at 120 Hz

3.6.5 Use with the QuadTech 1659 and 1692 Digibridges.

The 1659 and 1692 Digibridges have direct reading ranges well beyond the range of the 1417 standard but their specified accuracy gets very high above 1 F so that there is no point in checking accuracy above that value. The actual accuracy will be well below the values given in Table 3-16 if a SHORT calibration is made at the test frequency and many measurements are averaged. The 1659 applies 0.3 V max but the 1692 has a choice of 0.3 V or 1 V (open circuit). The 0.3V level of the 1692 should be used up to 1 mF at 120 Hz and up to 100 uF at 1 kHz. This 0.3 V level will cause an noticeable increase in the 100 uF value at 100 or 120 Hz, about 0.03%, which is less than the 1692 specification, but which should either be allowed for or reduced by reducing the signal level by putting about 30 ohms in series with the H CURRENT connection.

Table 3-16					
1659 and	1692	Specified	Accuracy		

1659				1692				
1417	100 &	120 Hz	1 k	1 kHz		: 120 Hz	1 kHz	
Setting	С	D	С	D	С	D	С	D
1 uF	.1%	.0005	.1%	.0005	.05%	.0003	.05%	.0003
10 uF	.1%	.0005	.105%	.00052	.05%	.0003	.055%	.00033
100 uF	.105%	.0005	.15%	.00075	.055%	.00033	.1%	.0006
1 mF	.15%	.00075	.6%	.0030	.10%	.0006	.55%	.0033
10 mF	.6%	.0030	5.1%	.0255	.55%	.0033	5.05%	.0303
100 mF	5.1%	.0255	-	-	5.05%	.0303	-	-
1 F	50%	.25	÷	-	50%	.3		-

3.6.6 Use with the QuadTech 1689, 1689M and 1693.

These instruments have direct reading capacitance ranges to 99999 uF (just under 100 mF) but this may be extended by using the RATIO mode. To do this:

a. Go to range 4 by measuring a high C value or setting

[4] [=] [SHIFT] [SPECIAL] [1]

b. Enter 1000 as the nominal value

[1000] [=] [NOM VAL]

c. Set RATIO mode

[2] [=] [SHIFT] [SPECIAL] [6] The reading is now in mF (although there is no unit displayed) and the range is extended to 99999 mF or 99.999 Farads.

To return to the normal mode enter [0] [=] [SHIFT] [SPECIAL] [6] These three instruments are very similar in behavior but are specified differently. They all have measurement accuracy at 1 kHz at 1 uF and 10 uF that is comparable to the 1417 ratio accuracy so that a true accuracy verification is difficult under these conditions (refer to paragraph 2.3.2), but these measurements can increase confidence. The 1417 ratio accuracy is adequate under most other conditions. At very high values the instrument specs are very broad and easily met. However, if a zero is made and many measurements averaged, the actual Digibridge accuracy can be quite good. 1% accuracy at 1 F at 100 Hz is possible.

These instruments have such tight tolerances that the changes in the 1417 values with signal level are important. This leads to a somewhat paradoxical situation in that lowering the signal level will give better values of the 1417 standard but will result in reduced specified accuracy of the instruments. Fortunately this occurs only at 100 or 120 Hz where the spec is broader. Moreover, the accuracy at reduced levels is much better than specified, particularily if several measurements are averaged, so that the specified accuracy at 1 V should be met and thus Table 3-17 gives these values even though it lists recommended signal levels that are substantially lower at 120 (or 100) Hz.

Table 3-17					
1689M	and	1693	Specified	Accuracy	

			1689M		1693				-	
1417	100 & 120 Hz		1 kHz		100 & 120 Hz		1 kHz		Test V	
Setting	С	D	С	D	С	D	С	D	120 Hz*	
1 uF	.04%	.0004	.02%	.0002	.04%	.0004	.02%	.0002	.25 V	
10 uF	.04%	.0004	.02%	.0002	.041%	.00041	.022%	.00022	.25 V	
100 uF	.04%	.0004	.05%	.0005	.047%	.00047	.04%	.0004	.1 V	
1 mF	.144%	.00144	.41%	.0041	.112%	.00112	.22%	.0022	.5 V	
10 mF	1.44%	.0144	4%	.04	.76%	.0076	2.02%	.0202	1 V	
100 mF**	14.4%	.144	-	-	7.24%	.0726	-	-	1 V	
1 F**	144%	1.4	-	-	72%	.72	-	-	1 V	

* Recommended programmed voltage at 100 or 120 Hz, see text ** Use RATIO mode to extend capacitance range

4.1 GENERAL.

-

Capacitors that are stable enough for use as standards would be extremely expensive at high capacitance values. Electrolytic types have poor stability. The 1417 is a network that simulates higher values, effectively multiplying the capacitance of its 1- μ F standard, a stabilized polystyrene capacitor, to higher values by means of two inductive voltage dividers (IVD's). It should not be used as a circuit component, for it cannot be used at dc and, generally, it must be used as a 4-terminal device (see para 2.8).

4.2 BASIC CIRCUIT.

The basic circuit of the GR 1417 is shown in Figure 4-1. C is the internal 1 μ F and the two autotransformers are IVD's with selectable ratios α and β . The open-circuit transfer impedance of this network $Z_T = E_0/I_{in}$. This transfer impedance is decreased by attenuating the input current by a factor α and the output voltage by β . If the dividers are ideal, that is they have exact and equal ratios, infinite opencircuit impedance and zero short circuit (winding) impedance, the equivalent for this network becomes a 2-terminal capacitor of value C/ $\alpha\beta$, where α and β are the nominal ratios.

Needless to say, the IVD's are not ideal, and each imperfection causes specific additions to such an ideal equivalent circuit.



Figure 4-1. Basic circuit diagram.

4.3 ERRORS IN EFFECTIVE 4-TERMINAL CAPACITANCE.

The transformers of Figure 4-1 are passive 3-terminal networks and can be replaced by equivalent, 3-impedance T networks, as shown in Figure 4-2. It should be noted that these impedances are slightly non-linear.

For this network the transfer admittance is

$$Y_{T} = \frac{1}{Z_{T}} = \frac{I_{in}}{E_{o}} = \left(\frac{Z_{2}+Z_{3}}{Z_{2}}\right) \left(\frac{Z_{5}+Z_{6}}{Z_{6}}\right) \times \left(j\omega C + \frac{1}{Z_{2}+Z_{3}} + \frac{1}{Z_{5}+Z_{6}}\right)$$
(1)

The first two factors are the reciprocals of the open-circuit divider ratios α and β in the 1417. It is relatively easy to get very precise ratios with high permeability toroidal cores. The dividers of the GR 1417 have ratio errors that are negligible compared to any of its specifications. The feature of the GR 1417 is that, if the last factor can be held constant, the ratio between capacitance settings depends only on these precise divider ratios. This is reflected in the specified ratio accuracy, which is the ratio between the capacitance at any setting and that at 1 μ F, if conditions are such that the last factor is constant.

The last factor is the admittance of capacitor C in parallel with the open-circuit input impedances of the dividers. These divider impedances are independent of the ratio settings α and β . If we assume that these divider impedances are purely inductive and (in parallel) have an inductance L, the last factor becomes:

$$j\omega C\left(1-\frac{1}{\omega^2 LC}\right) = j\omega C'$$
(2)

so that the transfer capacitance

$$C_{T} = \frac{Y_{T}}{j\omega} = \frac{C}{\alpha\beta} \left(1 - \frac{1}{\omega^{2}LC}\right) = \frac{C'}{\alpha\beta} \quad (3)$$

Where C' is the effective value of the standard C. This value has an error that depends on frequency squared, as shown. It also depends on signal level because of the non-linearity of inductance L. To keep this error small, the LC product should be kept as large as possible.

It was considered important to have the minimum value of capacitance (C') be 1 μ F, so that this value could be measured on precision bridges or compared against a precision standard (see para 2.4.). Available bridges and standards are substantially less accurate at higher values. This capacitance can be measured by a two-terminal bridge (A to A' in Figure 4-2).



Figure 4-2. General equivalent circuit.

13

Therefore, to keep the LC product high, the value of L is quite high, \approx 400 H. With a 1- μ F capacitor, this causes error of about 0.50% at 120 Hz. However, because L is reasonably constant, this error is compensated by additional padding capacitance at 100 Hz and 120 Hz, to give a direct-reading accuracy of 0.25% over a reasonable range of voltage levels. Correcting for this error can give much better accuracy (para 2.4) This error is negligible at 1 kHz.

There is an additional source of error which is important at higher frequencies. If there is mutual inductance between the input (CURRENT) circuit and the output (POTENTIAL) circuit, as shown in Figure 4-3a, it appears effectively as an inductance in series with the standard C_T (Figure 4-3b). This mutual inductance can be + or - and gives an error

$$C'_{T} = \frac{C_{T}}{1 \pm \omega^2 M C_{T}}$$

where C_T is the 4-terminal transfer capacitance. Usually the coupling internal to the 1417 is substantially less than the coupling in the external connections (para 2.2) but there is some. The effect of the internal coupling is included in the RATIO ACCURACY specification.



Figure 4-3. Equivalent circuits for mutual inductance errors.

4.4 TERMINAL IMPEDANCES.

Although a true 4-terminal measurement would be completely immune to impedance in series with any of the terminals, no instrument makes such an ideal measurement, so these impedances should be kept small to avoid measurement errors. The terminal impedances of the 1417 are shown in Figure 4-4. The lower two branch, or terminal, impedances of Figure 4-4 are small, for they are simply the impedance of the



Figure 4-4. Actual equivalent circuit.

wired common connection of Figure 4-1. The upper two terminal impedances are much more critical. From the equivalent circuit of Figure 4-2, the (open-circuit) input impedance can be shown to be:

$$Z_{in} = Z_1 + \frac{Z_2 Z_3}{Z_2 + Z_3} + \left(\frac{Z_2}{Z_2 + Z_3}\right)^2 \frac{1}{\left(i\omega C + \frac{1}{Z_2 + Z_3} + \frac{1}{Z_5 + Z_6}\right)}$$
(4)

which can be written as:

$$Z_{\rm in} = Z_{\rm isc} + \frac{a^2}{j\omega C'}$$
(5)

where Z_{isc} is the input impedance with a short across the capacitor. Subtracting the transfer impedance, we get the upper input terminal impedance

$$Z_{in} - Z_T = Z_{isc} + \frac{\alpha^2}{j\omega C'} - \frac{\alpha\beta}{j\omega C'} = Z_{isc} + \frac{\alpha(\alpha - \beta)}{j\omega C'}$$
(6)

(The other terminal impedance is easily found by symmetry.)

The first term is mostly resistive and a function of the winding resistance of the input divider. For a given core and winding area, this resistance is proportional to N^2 , where N is the number of turns. Because the open-circuit inductance is also proportional to N^2 , there is a tradeoff between keeping this resistance low and keeping the error of equation 2 small. The compromise reached was based on the characteristics of several instruments. These windings also have some leakage inductance, but the impedance of this inductance is small compared to the resistance, even at 1 kHz. These impedances are given in Table 4-2.

The second term in this terminal impedance is capacitive (+ or -) and is the result of inequality between the divider ratios α and β . These ratios can be exactly equal (nominally) when their product is 10^{-2} , 10^{-4} , or 10^{-6} , but, when it is 10^{-1} , 10^{-3} , or 10^{-5} , they can be equal only if they are irrational numbers. If these dividers are autotransformers and possible ratios are limited to ratios of integral numbers of turns, either these $\alpha\beta$ ratios must be slightly inaccurate or

 α and β must differ. A resistive divider across one turn of one divider can, however, give greatly increased resolution. Such a divider is used on the 10- μ F setting because this capacitive terminal impedance must be small to avoid an error on a specific bridge. Table 4-1 shows the winding scheme used and the resulting capacitive term of equation 6. Capacitance is always much less than the winding resistance, even at 50 Hz.

4.5 DISSIPATION FACTOR.

The dissipation factor of the 1417 is intentionally set to be .01 at 100 Hz, 120 Hz and 1 kHz (at the appropriate switch setting). This insures that, in spite of instrument errors, the D indication will be positive, allowing a valid D reading and proper operation. This is done by adding resistors in series with the standard capacitor, as shown in Figure 5-3.

The D value is affected by loss in the IVD's and changes with level slightly, but should remain within \pm .001. The D also includes loss in the standard capacitor and phase-shift errors in the IVD's but the effects are small.

Table 4-1 DIVIDER WINDING DATA

"CUR TURN	RENT" S a	"POTEN TURNS	TIAL" B	αβ ERROR	n(a-B)
3000	1	4000	1	0	×
949	.316333	1264%*	.316125	+9 ppm	15 mF
300	0.1	400	0.1	0	0*
96	.0320	125	.03125	0	42 mF
30	.01	40	.01	0	~
10	.003333	12	.003	0	.9 F
3	.001	4	.001	0	en
	"CUR TURN 3000 949 300 96 30 10 30	"CURRENT" TURNS n 3000 1 949 .316333 300 0.1 96 .0320 30 .01 10 .003333 3 .001	"CURRENT" "POTEN TURNS 3000 1 4000 949 .316333 1264%* 300 0.1 400 96 .0320 125 30 .01 40 10 .003333 12 3 .001 4	"CURRENT" TURNS "POTENTIAL" TURNS β 3000 1 4000 1 949 .316333 1264½* .316125 300 0.1 400 0.1 96 .0320 125 .03125 30 .01 40 .01 10 .003333 12 .003 3 .001 4 .001	$\begin{array}{c c} \mbox{``CURRENT''} & \mbox{``POTENTIAL''} & \mbox{σ} & \mbox{``TURNS} & \mbox{β} & \mbox{``ERROR} \\ \hline \mbox{3000} & 1 & 4000 & 1 & 0 \\ \hline \mbox{949} & .316333 & 1264\%^{\circ} & .316125 & +9 \mbox{ppm} \\ \hline \mbox{300} & 0.1 & 400 & 0.1 & 0 \\ \hline \mbox{96} & .0320 & 125 & .03125 & 0 \\ \hline \mbox{30} & .01 & 40 & .01 & 0 \\ \hline \mbox{10} & .003333 & 12 & .003 & 0 \\ \hline \mbox{3} & .001 & 4 & .001 & 0 \\ \hline \end{array}$

*Resistive divider across one turn

√ 10 = 3.16227766....

Table 4-2 TYPICAL TERMINAL IMPEDANCES

Setting	C	urrent	Potential				
	R	L	C*	R	L	C*	
1 µF	.019 Ω	.3 µH	00	.024 Ω	.54 µH	60	
10 µF	7.1 D	567 µH	15 mF	15.2 Ω	736 µH	-15 mF	
100 µF	3.06 Ω	123 µH	00	6.3 n	152	00	
1 mF	1.1 Ω	19.5 µH	42 mF	2.15 Ω	20.2	-43 mF	
10 mF	.36 Ω	3.5 µH	00	.70 Ω	4.3	00	
100 mF	.125 Ω	1.0 µH	.9F	.205 Ω	1.2	-1 F	
1 F	.035 Ω	.3 µH	00	.055 Ω	.4 µH	00	

*Calculated from ratios of Table 4-1.

5.1 GR FIELD SERVICE.

Our warranty attests the quality of materials and workmanship in our products. When difficulties do occur, our service engineers will assist in any way possible. If the difficulty cannot be eliminated by use of the following service instructions, please write or phone the nearest GR service facility, giving full information of the trouble and of steps taken to remedy it. Describe the instrument by type, serial, and ID numbers (Refer to front and rear panels.)

5.2 INSTRUMENT RETURN.

2

2

2

2

Before returning an instrument to GenRad for service, please ask our nearest office for a "Returned Material" number. Use of this number in correspondence and on a tag tied to the instrument will ensure proper handling and identification. After the initial warranty period, please avoid unnecessary delay by indicating how payment will be made, i.e., send a purchase-order number or (for transportation charges) request "C.O.D."

For return shipment, please use packaging that is adequate to protect the instrument from damage, i.e., equivalent to the original packaging. Advice may be obtained from any GR office.

5.3 CALIBRATION PROCEDURE.

If the low-level value of the 1 μ F CAPACITANCE setting for any TEST FREQUENCY position is found to be outside limits at 23° ± 5°C, (see section 2) its padding capacitors may be added or removed to bring it in tolerance. However, before repadding, cyclically demagnetize the divider cores as described in para 5.5, to ensure that the error is not caused by residual magnetization.

The padding capacitors for the 1-kHz position are on the capacitance standard C. This adjustment affects all TEST FREQUENCY switch positions and should be adjusted first. The 100-Hz and 120-Hz positions have separate padding capacitors and these padding capacitors are located on the TEST FREQUENCY switch, see Figure 5-1.

If the 1- μ F capacitance measurement indicates a short circuit, or if the waveform at the VOLTMETER terminals is non-sinusoidal (when a low-level sinusoidal signal is applied), the Zener diodes CR1 and CR2 may be shorted and should be replaced.

The other values of capacitance for the other CAPA-CITANCE settings are related to this $1-\mu F$ value by the process ratios of the inductive dividers T_1 and T_2 . Except for gross damage or burn out, these ratios should not change with time.



Figure 5-1. Internal view (transformer shields removed).

4 Ĩ E

5.4 DETERMINATION OF OTHER PARAMETERS.

5.4.1 Divider Ratios.

The divider ratios for both T_1 and T_2 can be checked by comparing them with a precision inductive voltage divider, using standard ratio measuring techniques. The input voltage (<20 V) should be applied at the VOLTMETER terminals and the outputs checked at both the CURRENT terminals and POTENTIAL terminals. The TEST FREQUENCY switch should be set to the EXTERNAL STANDARD position.

The ratios should be those values given in Table 4-1 to within $\pm.005\%$, from 50 Hz to 2 kHz. A high-resolution divider is required to measure the 1000 to 1 ratio to this precision. One should note that the 1417 could have applications as a precision divider.

5.4.2 Terminal Impedances.

The resistive and inductive parts of the terminal impedances can be measured with a suitable bridge. The GR 1608 is recommended because it can measure R and Q $(L = RQ/\omega)$. The measurements are made at the CURRENT terminals to measure Z_A (see specs) and at the POTENTIAL terminals to measure Z_B. The VOLTMETER terminals should be shorted. The parameters L and R vary only slightly with frequency.

The capacitive part of the terminal impedances is a function of the α and β ratios and can be determined by calculation. (See Section 4).

5.4.3 Divider Inductance.

The open-circuit inductance for the two dividers in parallel (L in equations of section 4) can be determined by measuring inductance across the VOLTMETER terminals with the TEST FREQUENCY switch in the EXTER-NAL STANDARD position. A low-frequency signal should be used because the dividers are resonant at 1 kHz (appear capacitive). The inductance should be approximately 400 H at 100 Hz, at low-levels.

5.5 SERVICE.

5.5.1 Demagnetization of Divider Cores.

Residual magnetization in the divider cores, resulting from the application of dc or a large ac transient, can be removed by cyclic demagnetization. In this procedure, the cores are cycled around their internal loops with a large signal, which is then slowly reduced in level to zero, leaving the cores unmagnetized. The amplitude of the signal should be less than 27 V or 2W, to avoid damage to the 39 V, 1.5-W Zener diodes that shunt the cores. Therefore, to get a high flux level, a very low frequency signal should be used; 20 V at 20 Hz is adequate and at lower frequencies the level may be proportionally lower. A GR 1310 set for full output at 5 Hz, can provide a satisfactory signal. Procedure:

 a. Set the 1417 TEST FREQUENCY switch to EXTER-NAL STANDARD.

 b. With the oscillator at low level, connect it to the 1417 VOLTMETER terminals.

c. Increase the oscillator level (not more than 25 V, see above).

d. Slowly decrease the oscillator level to zero.

5.5.2 Knob Removal.

If it should be necessary to remove the knob on a frontpanel control, either to replace one that has been damaged or to replace the associated control, proceed as follows:

a. Grasp the knob firmly with the fingers and pull the knob straight away from the panel.

b. Observe the position of the set screw in the bushing, with respect to any panel marking (or at the full ccw position of a continuous control).

c. Release the set screw and pull the bushing off the shaft.

d. Remove and retain the black Nylon thrust washer, behind the dial/knob assembly, as appropriate.

NOTE

To separate the bushing from the knob, if for any reason they should be combined off the instrument, drive a machine tap a turn or two into the bushing for a sufficient grip for easy separation.

5.5.3 Knob Installation.

To install a knob assembly on the control shaft:

a. Place the black Nylon thrust washer over the control
 shaft, if appropriate.

b. Mount the bushing on the shaft, using a small slotted piece of wrapping paper as a shim for adequate panel clearance.

c. Orient the set screw on the bushing with respect to the panel-marking index and lock the set screw.

NOTE

Make sure that the end of the shaft does not protrude through the bushing or the knob won't set properly.

 d. Place the knob on the bushing with the retention spring opposite the set screw.

e. Push the knob in until it bottoms and pull it slightly to check that the retention spring is seated in the groove in the bushing.

NOTE

If the retention spring in the knob comes loose, reinstall it in the interior notch with the small slit in the outer wall.

MECHANICAL PARTS LIST

1

2

3

1

2

3

2

2

2

2

2

2

2

1

2

2

1

7

FIG	ONT	DESCRIPTION	GR PART NO	FMC NO	MFGR PART NO
1	1	BINDING POST, CURRENT L, J4	0938-3000	24655	0938-3000
2	1	BINDING POST, CURRENT H, J3	0938-3000	24655	0938-3000
3	1	BINDING POST	0938-3022	24655	0938-3022
4	1	BINDING POST, POTENTIAL L, J2	0938-3000	24655	0938-3000
5	1	BINDING POST, POTENTIAL H, J1	0938-3000	24655	0938-3000
6	1	GASKET	5331-3100	24655	5331-3100
7	1	CABINET ASM	1417-1100	24655	1417-1100
8	4	FEET	5260-2060	24655	5260-2060
9	1	BINDING POST, VOLTMETER H, J5	0938-3000	24655	0938-3000
10	1	BINDING POST, VOLTMETER L. J6	0938-3000	24655	0938-3000
11	1	BINDING PCST	.0938-3022	24655	0938-3022
1Z	1	BINDING POST, EXTERNAL STANDARD H, J7	0938-3000	24655	0938-3000
13	1	BINDING POST, EXTERNAL STANDARD L. J8	0938-3000	24655	0938-3000
14	1	KNOB ASM, TEST FREQUENCY, S1 INCLUDES	5500-5421	24655	5500-5421
	1	RETAINER	5220-5401	24655	5220-5401
15	1	KNOB ASM, CAPACITANCE, SZ	5500-5421	24655	5500-5421
	1	RETAINER	5220-5401	24655	5220-5401



REFERENCE DESIGNATOR ABBREVIATONS

в	12	Motor
BT		Battery
С	*	Capacitor
CR		Diode
DS	.8	Lamp
F		Fuse
ł		Jack

		-
×	-	Gals

- = Relay = Relay Coil KL
- Relay Switch KS
- L . Inductor
- M - Meter
- MK - Microphone

- P = Plug
- Q = Transistor
- R = Resistor
- -S - = Switch
- τ - Transformer
- Integrated Circuit υ
- Diode, Zener VR
 - · Socket for Plug-In
- X Y - Crystal
- Network z

References ASA Y32.16 and MIL-STD-16C

ELECTRICAL PARTS LIST

CHASSIS MOUNTED PAPTS

RE	FDES	DESCRIPTION	PART NO.	FMC	MEGR PART NUMBER
с	100	CAPACITOR ASM 1UF .01PCT 160V	0236-4000	24655	0236-4000
С	102	CAP MYLAP . DOLUF 10 PCT 200V	4860-7309	56289	410P .001 UF 10PCT
С	103	CAP CER TUB 600PE 10PCT 50CV NM	4400-4900	00656	BC-CI-1-NM 600PF
С	200	CAP MYLAR .0047 UF SPCT 100V S	4860-7398	24655	4860-7398
C	201	CAP MYLAR .001UF 10 PCT 200V	4860-7309	56289	410P .001 UF 10PCT
С	202	CAP MYLAR . ODIUF 10 PCT 200V	4860-7309	56289	410P .001 UF 10PCT
С	273	CAP CER TUB 600PF 10PCT 500V NM	4400-4900	00656	BC-CI-1-NM 600PF
C	300	CAP MYLAR .0036 UF SPCT 200V S	4860-7374	56289	410P .0036 UF 5PCT
С	371	CAP MYLAR .001UF 10 PCT 200V	4860-7309	56289	410P .001 UF 10PCT
C	372	CAP MYLAP .001UF 10 PCT 200V	4860-7309	56289	410P .001 UF 10PCT
С.	3.73	CAP CER TUB SOOPE LOPCE SONV NM	4400-4900	00656	BC-CI-1-NM 600PF
CP	1	ZENER LMZ-394 39V 5PCT 1.5W	6083-1059	24444	LMZ-394 1.58398
CP	2	ZENER LM7-394 39V 50CT 1.5W	6083-1059	24444	LMZ-394 1.5R398
J	1	BINDING POST ASM	0938-3000	24655	0938-3000
J	2	BINDING POST ASM	0938-3000	24655	1938-3000
J	3	BINDING POST ASM	0938-3000	24655	0938-3000
J	4	BINDING POST ASM	0938-3000	24655	0938-3000
3	5	BINDING POST ASM	0938-3000	24655	0938-3000
J	6	BINDING POST ASM	0938-3000	24655	0938-3000
J	7	BINDING POST ASM	0938-3000	24655	0938-3000
1	8	BINDING POST ASM	0938-3000	24655	0938-3000
R	1	RES FLM 200HM 1/10PCT 50PPM 1/8W	6190-0004	75042	CEA 200 0HM 0.1PCT
R	2	RES FLM 12.7 0HM 1 PCT 1/8W	6250-9127	81349	RN55012R7F
R	3	RES FLM 2 OHM 1PCT 300PPM 1/2W	6619-2805	24655	6619-2805
R	4	RES COMP 6.8 OHM SPCT 1/4W	6099-9685	81349	RCR07G6R8J
R	5	RES FLM 2 OHM 1PCT 300PPM 1/2W	6619-2805	24655	6619-2805
R	6	RES FLM 2 OHM 1PCT 300PPM 1/2W	6619-2805	24655	6619-2805
2	7	RES COMP 62 DHM SPCT 1/4W D	6099-0625	81349	PCR076620.4
s	1	SWITCH ROTARY ASM	7890-5608	24655	7890-5608
S	2	SWITCH ROTARY ASM	-7890-5607	24655	7890-5607
т	1	TRANSFORMER, CURRENT DIVIDER	1417-4000	24655	1417-4000
T	2	TRANSFORMER POTENTIAL DIVIDER	1417-4010	24655	1417-4010

10 NOT PIGGY BACK PUT . (4) LEADS THRU TERMINALS FOR COMP. IN PARALLEL 7890-5608 Im 20 07 206 200 ci

SWITCH ASM (TEST FREQUENCY)

MAKE THE FOLLOWING CONNECTIONS

51	106 R	70	51 206R	BY	C 200.201,202,203 IN PARALLE
51	107R	70	SI. 207R .	BY	C300,301,302,303 IN PARALLER
51	IDZR	To	51. 202 F.R	BY	RI 5 ET IN PARALLEL
51,	103E	TO	SI, 203 F.R	BY	RZ
31,	104 R	To	51, 204 F.2	BY	23 5, R4 IN PARALLEL
51	1022	To	51, 103R TO	51.	104 2 BY ESMA1-30-20
51	206 R	70	51.207R		BY ESNW-30-20
51	207 R	70	51,208R 1	84	CR2 (CATHODE TO 20TR)
51,	208R	78	51, 201F. E	ry i	CRI (CATHODE TO 201R)

"IL

11

L.

N N

Figure 5-2. TEST FREQUENCY Switch (S2) details showing padding capacitors.



534 Main Street, Westbury, NY 11590 www.ietlabs.com TEL: (516) 334-5959 • (800) 899-8438 • FAX: (516) 334-5988